

# DIAGNOSTICS



*Your Key To Reliable Quality Controls*

## CERTIFICATE OF ANALYSIS

**MATERIAL** CYCLOSPORINE C2 CONTROL

**CATALOG NUMBER** 202

**LOT NUMBER** 5028

**STORAGE** LONG TERM STORAGE = <-14°C USE = 2 TO 8°C

**OUTDATE** = 2019-01-28 **MANUFACTURE DATE** = 2015-01-28

**INTENDED USE** For In Vitro Diagnostic Use

**PRECAUTION** ALL INGREDIENTS DERIVED FROM HUMAN SOURCE MATERIAL HAVE BEEN TESTED AND FOUND NEGATIVE FOR HEPATITIS B SURFACE ANTIGEN, HIV-1 & HIV-2 ANTIBODIES AND HEPATITIS C VIRUS ANTIBODIES. CERTIFICATES OF ANALYSIS FROM VENDORS ON FILE.

<u>SPECIFICATIONS</u>	<u>SPIKED TO LEVEL*</u>	<u>METHOD</u>	<u>RESULTS</u> MEAN/SD
LEVEL 1, LOT 50281 <u>VALUES</u> Cyclosporine	600 ng/mL	EMIT® 2000 Cyclosporine for C2 Monitoring	574.6/54.5 ng/mL
LEVEL 2, LOT 50282 <u>VALUES</u> Cyclosporine	1200 ng/mL	EMIT® 2000 Cyclosporine for C2 Monitoring	1261.5/81.5 ng/mL
LEVEL 3, LOT 50283 <u>VALUES</u> Cyclosporine	1800 ng/mL	EMIT® 2000 Cyclosporine for C2 Monitoring	1898.7/162.1 ng/mL

QC VERIFICATION *[Signature]* DATE 4-13-15

RA VERIFICATION *[Signature]* DATE 4-13-15

\* = Gravimetrically determined.